

**Headoffice:** C-4th&5th, 16, Deogyeong-daero 1556beon-gil, Yeongtong-gu, Suwon-si, Gyeonggi-do, 16690, REPUBLIC OF KOREA **Manufacturing site:** 14, Jeungpyeongsandan-ro, Jeungpyeong-eup, Jeungpyeong-gun, Chungcheongbuk-do, 27915, REPUBLIC OF KOREA

> TEL: +82-31-300-0400 / FAX: +82-31-300-0499 http://www.sdbiosensor.com

## **Certificate of Analysis**

Product Name	STANDARD™ M10 Arbovirus Panel			
Ref No.	M10-AB5-01	Cat No.	11ARB10A	
Lot No.	8C14981AJ1	MFG Date	2023.09.08.	
		EXP Date	2024.09.07.	
Testing categories	Specification			Result
Performance Inspection	Sensitivity: 100% detected in 2X LoD			Pass
	Repeatability: Results are within ± 2 Ct of target concentration			Pass
	Specificity: No fluorescence signal detected in negative control			Pass
Apperance Inspection	Check the text number of the labeling material Quantity per one container: 10ea			Pass
	Expiration date between cartridge label and package of test kit LOT No. between cartridge label and package of test kit			Pass

According to SD Biosensor's quality assurance system, the above QC tests have been completed and passed the specification. We hereby approve this product release.

SD Biosensor, Inc.

Ko Jongkwan

" उक्ष

Quality Management Representative QA Division

Ko Jongkwan

Quality Management Representative

QA Division Issue date: 2023-09-08

C-BD100-17 SD Biosensor, Inc.