

Headoffice: C-4th&5th, 16, Deogyeong-daero 1556beon-gil, Yeongtong-gu, Suwon-si, Gyeonggi-do, 16690, REPUBLIC OF KOREA **Manufacturing site:** 14, Jeungpyeongsandan-ro, Jeungpyeong-eup, Jeungpyeong-gun, Chungcheongbuk-do, 27915, REPUBLIC OF KOREA

 $\begin{tabular}{ll} TEL: +82-31-300-0400 / FAX: +82-31-300-0499 \\ & http://www.sdbiosensor.com \end{tabular}$

Issue date:

2023-08-08

Certificate of Analysis

Product Name	STANDARD™ M10 Sputum pretreatment Kit			
Ref No.	M10-SPTP-01	Cat No.	11PRT10A	
Lot No.	8R54881AJ1	MFG Date	2023.08.08.	
		EXP Date	2024.02.07.	
Testing categories	Specification		Result	
Performance Inspection	• Repeatability: Results are within ± 2 Ct of target concentration			Pass
	Specificity: No fluorescence signal detected in negative control			
	Check that there is no problem in detecting our products or applicable products			
Apperance Inspection	Check the text number of the labeling material Quantity of each product per one box: 10ea			Pass
	Expiration date of package of kitLOT No. of package of test kit			Pass

According to SD Biosensor's quality assurance system, the above QC tests have been completed and passed the specification. We hereby approve this product release.

SD Biosensor, Inc.

Ko Jongkwan

Quality Management Representative QA Division

Jongkwan Ko Quality Management Representative QA Division

C-BD100-17 SD Biosensor, Inc.