



Certificate of Analysis

Product Name	STANDARD™ Q RSV Ag Test		
Cat No.	09RSV40D	Ref No.	QRSV01G
Main Lot/Sub Lot	56367U1AC / 1	MFG Date	2025-07-28
		EXP Date	2027-06-25
Testing categories	Specification		Result
Apperance Inspection	<ul style="list-style-type: none">The mark on the appearance of the device is correct.Verify that the positions of the colored C line and T line are within the appropriate range. (Color Line Position Check)Buffer is a colorless liquid formulation.All desiccant color must be yellow.		Pass
Sensitivity	In-house reference panels for the sensitivity test : All panels must be interpreted as positive.		Pass
Specificity	In-house reference panels for the specificity test : All panels must be interpreted as negative.		Pass
Migration speed test	The migration of the product should appear in the results window within 2 minutes and should reach the end of the results window within 6 minutes.		Pass
Background condition test	Background of product should be light red or clean within the reading time.		Pass
Packaging Inspection	Test device (25ea) / Sterile swab (25ea) / Extraction buffer (25ea) / Nozzle cap (25ea) / Positive control swab (1ea) / Negative control swab (1ea) / IFU (1ea)		Pass
	Manufacturing and expiration date on the pouch and package of test kit.		Pass
	LOT No. on the pouch and package of test kit.		Pass

According to SD Biosensor's quality assurance system, the above QC tests have been completed and passed the specification. We hereby approve this product release.

SD Biosensor, Inc.

Ko Jongkwan

Quality Management Representative
QA Division

Jongkwan Ko
Quality Management Representative
QA Division