

# STANDARD F Dengue NS1 Ag FIA



## » Background

STANDARD F Dengue NS1 Ag FIA is a fluorescent immunoassay for the detection of Dengue virus NS1 antigen in human whole blood, serum, and plasma samples.

## » Product Specification




STANDARD F Dengue NS1 Ag FIA	
Intended use	Qualitative detection of Dengue virus NS1 antigen
Specimen type	Whole blood, Serum, Plasma
Specimen volume	100 µl
Testing time	5 - 15 mins
Storage condition	2 – 30°C / 36 - 86°F

## » Test Procedure



Insert the test device into the Test Slot of the analyzer.



Mix the 100µl of specimen (serum, plasma or whole blood) with assay diluent buffer.



Add the 100µl of mixed specimen.



Press the start button. The analyzer will automatically display the test result after 5 to 15 minutes.

## » Test Performance

Reference	Sensitivity	Specificity
RT-PCR	100% (130/130)	100% (280/280)

## » Ordering Information

Category	Product	Pack Size	CAT No.
Vector borne	STANDARD F Dengue NS1 Ag FIA	25 Tests	10DEN10D