

Declaration of Conformity



Manufacturer Name	SD Biosensor, Inc.	
Manufacturer Address	<u>Head Office</u> C-4th&5th, 16, Deogyong-daero, 1556beon-gil, Yeongtong-gu, Suwon-si, Gyeonggi-do 16690, KOREA <u>Manufacturing Site</u> 74, Osongsaengmyeong 4-ro, Osong-eup, Heungdeok-gu, Cheongju-si, Chungcheongbuk-do 28161, KOREA	
EC Representative Name	MT Promedt Consulting GmbH	
EC Representative Address	Altenhofstrasse 80 66386 St. Ingbert Germany	
Common Name	Immunoassay Test Kit	
Product Name	STANDARD™ F hs-CRP STANDARD™ F hs-CRP Control <i>*Please refer to "Annex I. Product List" on page 2 in more detail.</i>	
Reference Number	F-HCRP, C-F-HCRP	
Classification	Others not covered by Annex II and self-testing according to Directive 98/79/EC	
Conformity Assessment Route	Annex III of IVDD 98/79/EC (EC Declaration of Conformity)	
Applied Standards	EN ISO 13485:2016 EN ISO 18113-1:2011 EN ISO 14971:2012 EN ISO 18113-2:2011 EN ISO 23640:2015 EN ISO 15223-1:2016 EN ISO 17511:2003 EN 62366:2008 EN 13612:2002	

We herewith declare that the above mentioned products meet the provisions of the Council Directive 98/79/EC for in vitro diagnostic medical devices. All supporting documentations are retained under the premises of the manufacturer. SD Biosensor, Inc. is exclusively responsible for the declaration of conformity.

Place: Suwon-si, Republic of Korea
Valid from: May 13, 2022

Signature



Hyo-Keun, Lee
CEO / President

Annex I. Product List

F-HCRP

STANDARD™ F hs-CRP

- Test Device
(individually in a foil pouch with desiccant and Spoit™)
- Extraction buffer
- Instructions for use

EDMA Code
12 11 01 09 00

Description of EDMA code
C-Reactive Protein

C-F-HCRP

STANDARD™ F hs-CRP Control

- STANDARD F hs-CRP Control – Level 1
(10 tablets in the bottle, Transparent tube)
- STANDARD F hs-CRP Control – Level 2
(10 tablets in the bottle, Red colored tube)
- Instructions for use

EDMA Code
12 50 01 06 00

Description of EDMA code
Specific Protein Controls