



Headoffice: C-4th&5th, 16, Deogyong-daero 1556beon-gil, Yeongtong-gu, Suwon-si, Gyeonggi-do, 16690, REPUBLIC OF KOREA
Manufacturing site: 74, Osongsaengmyeong 4-ro, Osong-eup, Heungdeok-gu, Cheongju-si, Chungcheongbuk-do, 28161, REPUBLIC OF KOREA

TEL : +82-31-300-0400 / FAX : +82-31-300-0499
<http://www.sdbiosensor.com>

Certificate of Analysis

Product Name	STANDARD™ F RSV Ag FIA		
Cat No.	10RSV10D	REF No.	FRSV01G
Lot No.	64623X3AC	MFG Date	2021.03.31.
		EXP Date	2022.09.30.
Testing categories	Specification		Result
Apperance Inspection	<ul style="list-style-type: none"> • Uniform Gap between upper and lower device • No scratch, stain and crack etc. • No contamination 		<i>Pass</i>
Sensitivity	In-house reference panels for the sensitivity test 1) All panels must be interpreted as positive.		<i>Pass</i>
Specificity	In-house reference panels for the specificity test 1) The panel must be interpreted as negative.		<i>Pass</i>
Packaging Inspection	Strip Devices (25 strips) / Sterile swab (25ea) / Extraction buffer (25ea) / Filter cap(25ea)/Fixed volume dropper300µl(25ea)/Positive control swab(1ea)/Negative control swab(1ea)/ insert (1ea)		<i>Pass</i>

According to SD Biosensor's quality assurance system, the above QC tests have been completed and passed the specification. We hereby approve this product release.

SD Biosensor, Inc


Geunkuk Song
 Quality Management Representative
 QA Division

Geunkuk Song
 Quality Management Representative
 QA Division