

# Declaration of Conformity

<b>Manufacturer Name</b>	<b>SD Biosensor, Inc.</b>	
<b>Manufacturer Address</b>	<u>Head Office</u> C-4th&5th, 16, Deogyeong-daero, 1556beon-gil, Yeongtong-gu, Suwon-si, Gyeonggi-do 16690, KOREA	
	<u>Manufacturing Site</u> 74, Osongsaengmyeong 4-ro, Osong-eup, Heungdeok-gu, Cheongju-si, Chungchungbuk-do 28161, KOREA	
<b>EC Representative Name</b>	<b>MT Promedt Consulting GmbH</b>	
<b>EC Representative Address</b>	Altenhofstrasse 80 D-66386 St. Ingbert Germany	
<b>Common Name</b>	<b>Immunoassay Test Kit</b>	
<b>Model Name</b>	<b>STANDARD™ F Strep A Ag FIA</b> <i>*Please refer to "Annex I. Product List" on page 2 in more detail.</i>	
<b>Reference Number</b>	FSRA01G	
<b>Catalog Number</b>	10STR10B, 10STR10D	
<b>EDMA Code</b>	15.71.01.03	
<b>Classification</b>	<b>Others according to Annex II of Directive 98/79/EC</b>	
<b>Conformity Assessment Route</b>	Annex III of IVDD 98/79/EC (EC Declaration of Conformity)	
<b>Applied Standards</b>	EN ISO 13485:2012	EN ISO 18113-1:2011
	EN ISO 14971:2012	EN ISO 18113-2:2011
	EN ISO 15197:2015	EN ISO 15223-1:2016
	EN ISO 23640:2015	EN 62366:2008
	EN ISO 17511:2003	EN 13612:2002

We herewith declare that the above mentioned products meet the provisions of the Council Directive 98/79/EC for in vitro diagnostic medical devices. All supporting documentations are retained under the premises of the manufacturer. SD Biosensor, Inc. is exclusively responsible for the declaration of conformity.

**Place: Suwon-si, Republic of Korea**  
**Valid from: November 22, 2017**

Signature




---

**Hyo-Keun, Lee**  
**CEO / President**

## *Annex I. Product List*

**FSRA01G****STANDARD™ F Strep A Ag FIA**

- Test device
- Extraction buffer 1
- Extraction buffer 2
- Fixed volume dropper
- Sterile swab
- Positive control swab
- Negative control swab

**EDMA Code**

15.71.01.03