

# EU Declaration of Conformity



<b>Manufacturer (SRN)</b>	<b>SD Biosensor, Inc. (SRN : KR-MF-000009168)</b>
<b>Manufacturer Address</b>	<u>Head Office</u> C-4th&5th, 16, Deogyong-daero, 1556beon-gil, Yeongtong-gu, Suwon-si, Gyeonggi-do 16690, REPUBLIC OF KOREA  <u>Manufacturing Site</u> 14, Jeungpyeongsandan-ro, Jeungpyeong-eup, Jeungpyeong-gun, Chungcheongbuk-do 27915, REPUBLIC OF KOREA
<b>EC Representative (SRN)</b>	<b>MT Promedt Consulting GmbH (SRN: DE-AR-000000085)</b>
<b>EC Representative Address</b>	Ernst-Heckel-Straße 7 66386 St. Ingbert Germany
<b>Notified Body</b>	TÜV SÜD PRODUCT SERVICE GmbH
<b>Notified Body Address</b>	Ridlerstr. 65, 80339 München, Germany
<b>Notified Body No.</b>	0123
<b>Certificate No.</b>	EU Technical Documentation Assessment Certificate (IVDR): N/A EU QMS Certificate (IVDR): V13 075369 0072
<b>Product Name</b>	STANDARD™ M10 Hr-HPV
<b>Reference Number</b>	M10-HPV-02
<b>Catalogue Number</b>	11HPV20A
<b>Basic UDI-DI</b>	88001117IPCA01MA18XR
<b>Classification (Applied Rule)</b>	Class C (Rule 3(h) applied) according to Annex VIII from Regulation (EU) IVDR 2017/746
<b>Conformity Assessment Route</b>	Annex IX (Chapter I, II, III) of Regulation (EU) 2017/746 (IVDR)
<b>Common Specification</b>	Not Applicable

We herewith declare that the above mentioned products meet the provisions of the following Regulations/Directives:

- Regulation (EU) 2017 /746 for in vitro diagnostic medical devices

All supporting documentations are retained under the premises of the manufacturer. SD Biosensor, Inc. is exclusively responsible for the declaration of conformity.

**Place: Suwon-si, Republic of Korea**

**Valid from: November 21, 2025**

(Signed for and on behalf of SD Biosensor, Inc.)



**Signature**

**Name : Hyo-Keun, Lee**

**Position : CEO / President**