Importer Address: Plot No.38, Sector-4, IMT Manesar, Gurgaon, Haryana 122050, India
Headoffice: C-4th&5th, 16, Deogyeong-daero 1556beon-gil, Yeongtong-gu, Suwon-si, Gyeonggi-do, 16690, REPUBLIC OF KOREA
Manufacturing site: 14, Jeungpyeongsandan-ro, Jeungpyeong-eup, Jeungpyeong-gun, Chungcheongbuk-do, 27915, REPUBLIC OF KOREA
TEL : +82-31-300-0400 / FAX : +82-31-300-0499
http://www sdbiosensor.com

Certificate of Analysis

Product Name	STANDARD™ M10 MTB/NTM		
Ref No.	M10-MTNT-01	Cat No.	11MTB20A
Lot No.	8CC4971AJ1	MFG Date	2023.09.07.
		EXP Date	2024.09.06.

Testing categories	Specification	Result
Performance Inspection	• Sensitivity: 100% detected in 3X LoD	*Pass*
	• Repeatability: Target gene results are within ± 2 Ct of target concentration / Internal Control results are within ± 4 Ct of target concentration	*Pass*
	• Specificity: No fluorescence signal detected in negative control	*Pass*
Apperance Inspection	• Check the text number of the labeling material • Quantity per one container : 10ea	*Pass*
	• Expiration date between cartridge label and package of test kit • LOT No. between cartridge label and package of test kit	*Pass*

For overseas use only

According to SD Biosensor's quality assurance system, the above QC tests have been completed and passed the specification. We hereby approve this product release.

SD Biosensor, Inc.
Ko Jongkwan 고종관
Quality Management Representative
QA Division

Ko Jongkwan
Quality Management Representative
QA Division

Issue date: 2023-09-07