



Headoffice: C-4th&5th, 16, Deogyong-daero 1556beon-gil,
Yeongtong-gu, Suwon-si, Gyeonggi-do, 16690, REPUBLIC OF KOREA
Manufacturing site: 14, Jeungpyeongsandan-ro, Jeungpyeong-eup, Jeungpyeong-gun,
Chungcheongbuk-do, 27915, REPUBLIC OF KOREA

TEL : +82-31-300-0400 / FAX : +82-31-300-0499
<http://www.sdbiosensor.com>

F-QP-0026-0016-0163

Rev.00 (2025.01.18.)

Certificate of Analysis

Product Name	STANDARD M10 MTB-RIF/INH		
Cat No.	11MTB30A	Ref No.	M10-MTB-02
Lot No.	8D574Q2AJ2	MFG Date	2026-04-24
		EXP Date	2027-04-23
Testing categories	Specification		Result
Performance Inspection	• Sensitivity: 100% detected in concentration criteria		Pass
	• Repeatability: Results are within ± 2 Ct of target concentration		Pass
	• Specificity: No fluorescence signal detected in negative control		Pass
Apperance Inspection	• Check the text number of the labeling material • Check component quantity		Pass
	• Check Lot No., MFG date, EXP date of all components		Pass
	• Leakage Inspection - Check the sealing of the packaging pouch		Pass

According to SD Biosensor's quality assurance system, the above QC tests have been completed and passed the specification. We hereby approve this product release.

SD Biosensor, Inc.

Ko Jongkwan

Quality Management Representative
QA Division

Jongkwan Ko
Quality Management Representative
QA Division