



Certificate of Analysis

Product Name	STANDARD™ M10 STI Panel		
Cat No.	11STI10A	Ref No.	M10-STI-01
Main Lot/Sub Lot	8CM5C61AJ1	MFG Date	2024-12-06
		EXP Date	2025-12-05
Testing categories	Specification		Result
Performance Inspection	• Sensitivity: 100% detected in 1,000 copies/ml		Pass
	• Repeatability: Results are within ± 2 Ct of target concentration		Pass
	• Specificity: No fluorescence signal detected in negative control		Pass
Apperance Inspection	• Check the text number of the labeling material		Pass
	• Quantity per one container : 10ea		
	• Expiration date between cartridge label and package of test kit		Pass
	• LOT No. between cartridge label and package of test kit		

According to SD Biosensor's quality assurance system, the above QC tests have been completed and passed the specification. We hereby approve this product release.

SD Biosensor, Inc.

Ko Jongkwan

Quality Management Representative
QA Division

Jongkwan Ko
Quality Management Representative
QA Division