



Headoffice: C-4th&5th, 16, Deogyong-daero 1556beon-gil,
Yeongtong-gu, Suwon-si, Gyeonggi-do, 16690, REPUBLIC OF KOREA
Manufacturing site: 74, Osongsaengmyeong 4-ro, Osong-eup, Heungdeok-gu,
Cheongju-si, Chungcheongbuk-do, 28161, REPUBLIC OF KOREA

TEL : +82-31-300-0400 / FAX : +82-31-300-0499
<http://www.sdbiosensor.com>

Certificate of Analysis

| Product Name | STANDARD™ F RSV Ag FIA | | |
|-----------------------|--|----------|-------------|
| Cat No. | 10RSV10D | Ref No. | FRSV01G |
| LOT No. | 64639U2AC | MFG Date | 2022.09.28. |
| | | EXP Date | 2024.03.27. |
| Testing categories | Specification | | Result |
| Appearance Inspection | <ul style="list-style-type: none"> • Uniform Gap between upper and lower device • No scratch, stain and crack etc. • No contamination • There should be colored lines. • Check device type • Check buffer amount | | Pass |
| Barcode Check | The product name and 'apply sample' should be displayed on the screen. | | Pass |
| Sensitivity | In-house reference panels for the sensitivity test 1) All panels must be interpreted as positive. | | Pass |
| Specificity | In-house reference panels for the specificity test 1) All panels must be interpreted as negative. | | Pass |
| Packaging Inspection | Test Device (25ea) / Sterile swab (25ea) / Extraction buffer (25ea) / Nozzle cap (25ea) / Fixed volume dropper(300ul) (25ea) / Positive control swab (1ea) / Negative control swab (1ea) / IFU (1ea) | | Pass |
| | No leakage in sealing part of pouch | | Pass |
| | No leakage in sealing part of buffer solution tube | | Pass |
| | Manufacturing and expiration date between the pouch and package of test kit. | | Pass |
| | LOT No. between the pouch and package of test kit. | | Pass |

According to SD Biosensor's quality assurance system, the above QC tests have been completed and passed the specification. We hereby approve this product release.

SD Biosensor, Inc.

Ko Jongkwan

Quality Management Representative
QA Division

Ko Jongkwan
Quality Management Representative
QA Division