

Headoffice: C-4th&5th, 16, Deogyeong-daero 1556beon-gil, Yeongtong-gu, Suwon-si, Gyeonggi-do, 16690, REPUBLIC OF KOREA Manufacturing site: 74, Osongsaengmyeong 4-ro, Osong-eup, Heungdeok-gu, Cheongju-si, Chungcheongbuk-do, 28161, REPUBLIC OF KOREA / 18-29, Cheomdangieop 2-ro, Sandong-eup, Gumi-si, Gyeongsangbuk-do, 39174, Republic of Korea TEL : +82-31-300-0400 / FAX : +82-31-300-0499 http://www.sdbiosensor.com

Certificate of Analysis

Product Name	STANDARD™ M10 Flu/RSV/SARS-CoV-2			
Ref No.	M10-CVFR-01	Cat No.	11FLU10A	
Lot No.	8CB42 6 AJ1	MFG Date	2023.02.06	
		EXP Date	2024.02.05	
Testing categories	Specification			Result
Performance Inspection	Sensitivity: 100% detected in 2X LoD			Pass
	• Repeatability: Results are within ± 2 Ct of target concentration			Pass
	Specificity: No fluorescence signal detected in negative control			Pass
Apperance Inspection	Check the text number of the labeling materialQuantity per one container : 10ea			Pass
	 Expiration date between cartridge label and package of test kit LOT No. between cartridge label and package of test kit 			Pass

According to SD Biosensor's quality assurance system, the above QC tests have been completed and passed the specification. We hereby approve this product release.

SD Biosensor, Inc.



Quality Management Representative QA Division

Ko Jongkwan

Jongkwan Ko Quality Management Representative QA Division

Issue date: 2023.02.06